



NOTES: (UNLESS OTHERWISE SPECIFIED)

1. THIS IS A CRYOGENIC VACUUM COMPONENT.
2. WELDING PROCEDURE: PER VENDOR SPECIFICATION WITH LBNL APPROVAL.
3. CLEANING PROCEDURE : PER VENDOR SPECIFICATION WITH LBNL APPROVAL.
4. PACKAGING AND STORAGE PROCEDURE OF THE COMPONENTS: PER VENDOR SPECIFICATION WITH LBNL APPROVAL.
5. DIMENSIONS AND TOLERANCING PER ANSI Y14.5M-1982. UNITS ARE IN INCHES [mm] UNLESS OTHERWISE SPECIFIED.
6. USE OF SULFUR OR SILICONE BEARING OILS, LUBRICANTS, OR COOLANTS ARE STRICTLY PROHIBITED.
7. USE OF RESIN OR RUBBER BONDED ABRASIVES UNDER POWER IS STRICTLY PROHIBITED. USE VITREOUS BONDED ABRASIVES ONLY.
8. VENDOR SUGGESTED CHANGES TO WELD PREPS; SUBJECT TO LBNL APPROVAL.
9. FITTINGS MAY BE USED IN PLACE OF BENDS; SUBJECT TO LBNL APPROVAL.
10. VENDOR SUGGESTED CHANGES TO TOLERANCES TO FACILITATE FABRICATION OR ASSEMBLY; SUBJECT TO LBNL APPROVAL.
11. REMOVE ALL THE BURRS AND REAM THE ENDS FOR CIRCULARITY AND CLEAN ENDS.
12. TUBE END SURFACE MUST BE PERPENDICULAR TO THE TUBE AXIS WITHIN +/- .010.
13. PERFORM ACCEPTANCE TESTS PER LBNL SPECIFICATION M989.
14. A MARK DESIGNATING THE INSTALLED LENGTH WILL BE UTILIZED DURING FINAL INSTALLATION OF THE FEEDBOX ASSEMBLY. MARK, SCRIBE OR ETCH THIS LOCATION IN A PERMANENT MANNER, SUBJECT TO LBNL APPROVAL, TO AN ACCURACY OF ± 0.063 ".
15. PROVIDE A MINIMUM LENGTH OF 4.0" OF STRAIGHT, SMOOTH PIPE ON THE INDICATED SIDE OF THE INSTALLED LENGTH MARK FOR PIPE WELDING DURING FINAL INSTALLATION OF THE FEEDBOX ASSEMBLY.
16. PIPE MUST BE STRAIGHT AND SMOOTH (NO BUMPS) FOR 1.5" ON EITHER SIDE OF THE CENTER-PLANE OF THE SUPPORT.
17. CAP BOTH ENDS OF PIPE TO FACILITATE ACCEPTANCE TESTS.

SCALE 9/32

THIRD ANGLE PROJECTION

6	-	I	FLANGE, SPIDER INTERFACE	SS304L
5	-	I	TUBE, PER ASTM A269	SS304L
4	-	I	TUBE, PER ASTM A269	SS 304L
3	-	I	TUBE, PER ASTM A269	SS 304L
2	-	I	BRAIDED FLEX HOSE, 1/2" ID X 3" LL	SS-300 SERIES
1	-	I	BRAIDED FLEX HOSE, 1/2" ID X 3" LL	SS-300 SERIES
ITEM	PART NO.	REQD	DESCRIPTION	MATERIAL
1			FLANGE, SPIDER INTERFACE	SS304L
2			TUBE, PER ASTM A269	SS304L
3			TUBE, PER ASTM A269	SS 304L
4			TUBE, PER ASTM A269	SS 304L
5			BRAIDED FLEX HOSE, 1/2" ID X 3" LL	SS-300 SERIES
6			BRAIDED FLEX HOSE, 1/2" ID X 3" LL	SS-300 SERIES
ERNEST ORLANDO LAWRENCE BERKELEY NATIONAL LABORATORY UNIVERSITY OF CALIFORNIA - BERKELEY				
LHC IR FEEDBOX CRYOGENICS PIPE, CC3, DFBX D8H				
MICROFILMED:		DWG. TYPE	SHOWN ON	SCALE: 9/32
ASSEM		DATE: 13-Dec-01	PATENT CLEAR:	DESIGN ACCT. NO.
ZSLCE2		DATE: 04-Nov-02	DATE: 04-Nov-02	DATE: 04-Nov-02
LH2003		DATE: 04-Nov-02	DATE: 04-Nov-02	DATE: 04-Nov-02
2512386		DATE: 04-Nov-02	DATE: 04-Nov-02	DATE: 04-Nov-02
1		DATE: 04-Nov-02	DATE: 04-Nov-02	DATE: 04-Nov-02

2512386

SHEET 1 OF 1